

Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit

FOR IN VITRO DIAGNOSTIC USE ONLY.
FOR SELF-TESTING.

PLEASE READ INSTRUCTIONS CAREFULLY BEFORE YOU PERFORM THE TEST.



Scan to watch the operation video and IFU of this test kit

Test Kit contents



Test kit contains test cassettes, sterile anterior nasal swabs, sampling tubes containing individual lysis buffer, droppers, and instructions for use.

REF	Specification	Test Cassette	Sterile swabs	Lysis Buffer and Dropper	Bio-Safety Bag	Instructions For Use
123143-01-102S	1pc/box	1	1	1	1	1
123143-05-102S	5pcs/box	5	5	5	5	1

Materials required but not provided: timer

Introduction

Coronavirus (CoV) belongs to the order Nidovirales under the Coronaviridae family with 4 genera: α , β , γ and δ . The α and β genera are only pathogenic to mammals, while γ and δ genera mainly cause bird infections. CoV is mainly transmitted through direct contact with secretions or through aerosols and droplets. There is also evidence supporting fecal-oral transmission.

7 kinds of human coronaviruses (HCoV) that cause human respiratory diseases have been identified so far, including: HCoV-229E, HCoV-NL63, HCoV-OC43, HCoV-HKU1, SARS-CoV, MERS-CoV and SARS-CoV-2. SARS-CoV-2 is one of the most contagious viral pathogens that causes human respiratory tract infections (RTI). Currently, the patients infected by SARS-CoV-2 are the main source of infection; asymptomatic infected people can also be an infectious source. Based on the current epidemiological investigation, the incubation period is 1 to 14 days, mostly 3 to 7 days. The clinical manifestations include fever, fatigue, cough and other symptoms, accompanied by dyspnea, which can rapidly develop into life-threatening severe pneumonia, respiratory failure, acute respiratory vesicle syndrome, septic shock, multiple organ failure, and severe metabolic acid-base imbalance.

Influenza, usually called flu, is an acute respiratory infection caused by Influenza virus. It is highly contagious. It is mainly spread through coughing and sneezing. It usually breaks out in spring and winter. Influenza A viruses and influenza B viruses are thought to be the virus types causing epidemics. Influenza A viruses are highly variable, followed by Influenza B viruses. Therefore, Influenza A viruses are more prevalent and severe, followed by Influenza B viruses. Influenza A includes H1N1, H3N2, H5N1, H7N9, and Influenza B includes Influenza B (Victoria) and Influenza B (Yamagata).

Human adenovirus (ADV) belongs to the adenoviridae family, mammalian adenovirus genus, which is a double-stranded DNA virus without an envelope, mainly infects human respiratory tract, digestive tract and urogenital tract. The main ADV related to respiratory diseases is ADV-B Group (ADV-3, 7, 11, 14, 16, 21, 50, 55), ADV-C Group (ADV-1,2,5,6) and ADV-E group (ADV-4). Acute respiratory adenovirus (ADV) infection which is one of the most common acute respiratory infections in infants and young children. It mainly causes fever, cough, dyspnea and other symptoms.

Respiratory syncytial virus (RSV) belongs to Pneumovirus of Paramyxoviridae with only one serotype, which is a single stranded negative-strand RNA virus with an envelope. RSV infection mainly causes bronchiolitis and pneumonia in infants under 6 months of age and upper respiratory tract infections such as rhinitis and colds in older children and adults.

Intended Use

Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit is an immunochromatography based one step in vitro test. It is designed for the qualitative detection of the SARS-CoV-2 virus, Influenza A virus, Influenza B virus, Adenovirus and Respiratory syncytial virus in human anterior nasal swab samples. The test results are used as an aid for diagnosis of respiratory pathogen infections, and are suitable for people with clinical symptoms such as fever, sore throat, cough, runny nose. The test kit is designed for use as self-testing. It is designed to detect SARS-CoV-2 within the first 7 days of symptom onset, and Influenza A/B, ADV or RSV within the first 4 days of symptom onset.

This test kit is not used in combination with other equipment and is not automated.

Test Principle

Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit uses a double antibody sandwich method to detect SARS-CoV-2, Influenza A, Influenza B virus, Adenovirus (ADV), Respiratory syncytial virus (RSV) by colloidal gold immunochromatography. When the appropriate amount of test samples treated with lysis buffer is added to the sample well of the test cassette, the sample will move forward along the test strip by capillary action. If the sample contains SARS-CoV-2, Influenza A, Influenza B, Adenovirus (ADV), Respiratory syncytial virus (RSV) antigen, and the concentration is higher than the limit of detection, the antigen will form immune complexes with corresponding Nucleocapsid Protein antibody labeled with colloidal gold respectively, which are captured by lines nCoV line, Flu A line, Flu B line, ADV line, RSV line. If test sample contains SARS-CoV-2 virus, forming a red nCoV line, indicating a positive result for SARS-CoV-2. If test sample contains Influenza A virus, forming a red Flu A line, indicating a positive result for Influenza A. If test sample contains Influenza B virus, forming a red FluB line, indicating a positive result for Influenza B. If test sample contains Adenovirus, forming a red ADV line, indicating a positive result for Adenovirus. If test sample contains Respiratory syncytial virus, forming a red RSV line, indicating a positive result for Respiratory syncytial virus. The C line should be formed to indicate that the sample has been transported properly through the membrane regardless of whether sample contains antigens or not. If the C line does not appear, it indicates that the test result is invalid and the sample need to retest.

Storage Instructions

Validity period: 36 months at 2°C to 30°C. Do not use after expiry. Do not freeze. This test kit should be used within 1 hour after opening the foil pouch.

Testing Procedure

Note:

1. Materials required but not included: timer.

2. This test kit can be used independently by individuals who are 18 or older. For those under the age of 18, it should be operated or supervised by an adult.

- Bring the kit to room temperature before testing. Wash your hands and dry.
- Check the expiry date on the box or foil bag. Check that the Test kit contents have not been previously used (these disposable materials cannot be used twice). Do not open pouch until ready to use.
- Tear the seal of the sampling tube and put it into the test-tube rack.
- Collecting the sample: Open the package containing the sterile swab. Avoid touching the cotton tip and remove the swab using the plastic handle.
- Insert the swab into a nostril (2.5 cm). Be sure to collect any nasal drainage that may be present. Carefully rotate the swab in a circular path against the inside of the nostril at least 5 times. Using the same swab repeat the procedure in the other nostril.
- After collecting the sample, insert the swab's cotton tip into sampling tubes containing individual lysis buffer, rotate the swab against the inner tube wall 10 times.
- Squeeze the swab from the outer tube wall 5 times to completely dissolve the sample in the buffer.
- Move the swab up until it is resting on the sample solution, squeeze the swab from the outer tube wall in order to leave the sample in the tube as much as possible.
- Remove and discard the swab, cover the tube with the dropper.
- Dispense 4 drops (approximately 100 µL) into each of the 2 sample wells in the test cassette. Read the results after 15 minutes. Do not read after 20 minutes. Note: If the dispensed drop contains air bubbles, add another drop into the well.
- Dispose all those used materials into Bio-safety bag and seal well. Then bio-safety bag can be disposed in a household bin.

Display of Results/Expected Values

Positive result:

SARS-CoV-2 positive result: If both the control line (C line) and the test line (nCov line) appear at the same time, it means that SARS-CoV-2 antigen has been detected and the result is positive.

There is currently a suspicion of a COVID-19 infection.

• If you test positive, you should not visit high-risk settings like hospitals and aged and disability care settings for at least 7 days or until symptoms have gone, unless seeking immediate medical care.

• To help protect those around you, we recommend to avoiding contact with people who are at higher risk of severe disease, wearing a mask outside the home, working from home where possible, avoiding going to school, public areas, or travel on public transport, in taxis or ride-share services, practicing good hygiene, and following your local health department's advice when leaving home.

• If you have any appointments you cannot miss (visit to a doctor, family violence service or police), let them know in advance that you have COVID-19.

• If you have a COVID-19 positive result, staying at home protects the people in your community and you should not visit high-risk settings like hospitals and aged and disability care settings.

• If you feel unwell or need COVID-19 advice for someone in your care, talk with your health provider, or speak to a nurse by calling the health direct helpline on 1800 022 222.

• If you develop symptoms such as severe shortness of breath or chest pain, call triple zero (000) immediately. Tell the call handler and the paramedics on arrival if you have COVID-19.

• Most people who test positive for COVID-19 recover completely, but some people may develop long COVID. COVID-19 vaccinations, including boosters, reduce your risk of re-infection and gives the best protection against severe illness from COVID-19. After testing positive, you should wait 6 months before making a booster dose appointment.

Influenza A positive result: If both the control line (C line) and the Influenza A test line (Flu A line) appear at the same time, it means that Influenza A antigen has been detected in the sample and the result of Influenza A is positive.

Influenza B positive result: If both the control line (C line) and the Influenza B test line (Flu B line) appear at the same time, it means that Influenza B antigen has been detected in the sample and the result of Influenza B is positive.

ADV positive result: If both the control line (C line) and the test line (ADV line) appear at the same time, it means that ADV antigen has been detected in the sample and the result of ADV is positive.

RSV positive result: If both the control line (C line) and the test line (RSV line) appear at the same time, it means that RSV antigen has been detected in the sample and the result of RSV is positive.

If the two quality control C lines appear, and more red lines appear in the detection line area, indicating that the sample contains one or more pathogenic microorganisms.

Negative result: If only the quality control C line appears and the detection line is not visible, the sample contains no SARS-CoV-2, Influenza A, Influenza B, ADV, RSV or the concentration is lower than the limit of detection and the result is negative.

Invalid result: If the C line does not appear, the result is invalid and a new test must be performed again.

Note:

The intensity of color that the test line area (nCov line/Flu A line/Flu B line) shows will vary according to the concentration of SARS-CoV-2 antigen, Influenza A antigen and Influenza B antigen. The result should be determined on whether the test line is formed or not, and is irrelevant to the color intensity. Therefore, any intensity of color in the test area (nCov line/Flu A line/Flu B line) should be considered positive.

The intensity of color that the test line area (ADV line/RSV line) shows will vary according to the concentration of ADV antigen, RSV antigen antigen. The result should be determined on whether the test line is formed or not, and is irrelevant to the color intensity. Therefore, any intensity of color in the test area (ADV line/RSV line) should be considered positive.

For SARS-CoV-2&Flu A/B

"C": Control Line "FluA": Influenza A Test Line "FluB": Influenza B Test Line
"nCov": SARS-CoV-2 Test Line "S": Sample Well

If you get a positive result, there is currently a suspicion of respiratory tract infection, individuals with a positive result or who are unwell are advised to consult a medical practitioner for follow up clinical care.

A negative result does not mean a person is not with respiratory tract infection. If symptoms persist the person should seek medical attention and further testing by PCR if required.

If an invalid result is produced, the user should retest with a new test.

For ADV/RSV

"C": Control Line "A": ADV Test Line "R": RSV Test Line "S": Sample Well

If you get a positive result, there is currently a suspicion of respiratory tract infection, individuals with a positive result or who are unwell are advised to consult a medical practitioner for follow up clinical care.

A negative result does not mean a person is not with respiratory tract infection. If symptoms persist the person should seek medical attention and further testing by PCR if required.

If an invalid result is produced, the user should retest with a new test.

Test Method Limitations

- 1. The accuracy of the test is dependent on the quality of the sample. Improper sampling or storage, using expired samples or repeated frozen-thawed samples can affect the test results. Test results can also be affected by temperature and humidity.
- 2. Negative results may be caused by low concentration of SARS-CoV-2, Influenza A, Influenza B, ADV, RSV antigens in the sample and therefore cannot completely rule out the possibility of infection.
- 3. False Negative results may be caused by low concentration of SARS-CoV-2, Influenza A, Influenza B, ADV, RSV antigens in the sample or Some use errors can also lead to false negative, therefore cannot completely rule out the possibility of infection.
- 4. Some medication (e.g. high concentration of over-the-counter (OTC) or prescription medication such as nasal spray) in the collected samples may interfere with the test result. Please perform the test again if the result is in doubt.
- 5. This product is only for qualitative testing and the specific concentration of each indicator must be measured using other quantitative methodologies.
- 6. The test results of this kit are for clinical reference only and are not the sole basis for clinical diagnosis. The clinical diagnosis and treatment of patients should be comprehensively considered in combination with their symptoms/signs, medical history, other laboratory tests and treatment response.
- 7. This product is less reliable in the later phase of infection and in asymptomatic individuals.
- 8. Repeat testing (within 1-3 days) is recommended in high-risk settings or if there is an ongoing suspicion of infection.
- 9. A positive result from this product does not necessarily determine whether a person is infectious.
- 10.This test kit can be used independently by individuals who are 18 or older. For those under the age of 18, it should be operated or supervised by an adult.
- 11. Test kit solutions should only be used as directed; do not ingest; do not dip the swab into provided solution or other liquid before inserting the swab into the nose; avoid contact with skin and eyes; keep out of the reach of children and pets before and after use. If the extraction buffer comes in contact with the skin or eyes, flush with plenty of water. If irritation persists, seek medical advice from a doctor or your local medical centre.
- 12. If the SARS-CoV-2 test result is positive, there is currently a suspicion of a COVID-19 infection, Follow the guidance from your local State or Territory Health Department for guidance on confirmation testing if necessary, and if unwell seek medical assistance.
- 13. If influenza A/B test result is positive: There is currently a suspicion of influenza A/B infection, individuals with a positive result or who are unwell are advised to consult a medical practitioner for follow up clinical care.
- 14. Do not re-use the test kit.
- 15. Do not use the test kit if the pouch is damaged, the seal is broken or the test cassette is wet or polluted.
- 16. Do not use the test kit contents beyond the expiration date printed on the outside of the box.
- 17. When collecting an anterior nasal swab sample, use only the Anterior Nasal Swab provided in the Kit.
- 18. If an invalid result is produced, the user should retest with a new test.
- 19. Do not mix with kit components from other batches.

Product Performance

• **Limit of Detection-LoD**
Limit of Detection (LoD) studies determined the lowest detectable concentration of SARS-CoV-2, Influenza A, Influenza B , ADV, RSV at which 100% of all (true positive) replicates test positive. (TCID₅₀/ml: The Common units of virulence of live viruses)

	Virus Strain	LoD (TCID ₅₀ /ml)		Virus Strain	LoD(TCID ₅₀ /mL)
SARS-CoV-2	BetaCoV/J512/human/2022	10	ADV	ADV-1	10 ⁵
	A/Brisbane/02/2018 (H1N1)	10 ⁴		ADV-2	10 ³
Influenza A	A/PUERTO/8/1934 (H1N1)	10 ²		ADV-3	10 ⁴
	A/Kansas/14/2017 (H3N2)	10 ²		ADV-4	10 ⁴
	A/Aichi/2/1968 (H3N2)	10 ²		ADV-7	10 ³
	A/Anhui/1/2013 (H7N9)	10 ⁴		ADV-55	10 ³
	B/Colorado/06/2017 (Victoria)	10 ⁰	RSV	RSV-A	10 ⁴
Influenza B	B/Phuket/3073/2013 (Yamagata)	10 ²		RSV-B	10 ⁴
	B/Chaoyang Beijing/12977/2017 (Yamagata)	10 ⁴			

Analytical Inclusivity
The folowing variants can be detected with the Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit.
SARS-CoV-2:Alpha, Beta, Gamma, Delta, Omicron, XBB.1.5,XBB.1.16.
FluA:A/Victoria/4897/2022 (H1N1), A/Croatia/10136RV/2023 (H3N2), A/Wisconsin/67/2022 (H1N1), A/District of Columbia/27/2023 (H3N2), A/Thailand/8/2022 (H3N2), A/Massachusetts/18/2022 (H3N2), A/Sydney/5/2021 (H1N1), A/Darwin/9/2021 (H3N2), A/Darwin/6/2021 (H3N2), A/Victoria/2570/2019 (H1N1), A/Wisconsin/588/2019 (H1N1), A/Hong Kong/2671/2019 (H3N2), A/Hong Kong/45/2019 (H3N2).
FluB:B/Austria/1359417/2021, B/Phuket/3073/2013, B/Washington/02/2019, B/Colorado/06/2017, B/Victoria/27/2020,B /Brisbane/60/2008.
ADV: ADV-1,2,3,4,7,55.
RSV:RSV Type A/A2, RSV Type A/Long, RSV Type B/GZ/1704-8, RSV Type B/18537, RSV Type B/B WV-14617-85.

• **Cross Reactivity**
Cross reactivity of Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit was evaluated by testing commensall and pathogenic microorganisms listed below that may be present in the clinical samples. The substance below are no Cross Reactivity with test result,which the each assay marker is active on itself and does not cross-react with the other marker(For example, the SARS-CoV-2 is active against SARS-CoV-2, but has no cross-reactivity against H1N1 (2009), Influenza A H1N1, Seasonal Influenza A H3N2, Influenza A H5N1, etc).
List of cross reactivity sustances tested:(The test concentration is 1×10⁶ CFU/mL or 1×10⁶ TCID₅₀/mL). Human coronavirus OC43, Human coronavirus NL63, Human coronavirus HKU1, Human coronavirus 229E, MERS-coronavirus, SARS-coronavirus, SARS-CoV-2, H1N1(2009), Influenza A H1N1, Seasonal Influenza A H3N2, Influenza A H5N1, Influenza A H7N9, Influenza B Victoria, Influenza B Yamagata, Parainuenza virus Type 1,Respiratory syncytial virus,Enterovirus CA16e, ADV-1, ADV-2, ADV-3, ADV-4, ADV-7, ADV-55, RSV-A, RSV-B, Mycoplasma pneumoniae, Staphylococcus aureus, Staphylococcus epidermidis, Bordetella pertussis, Legionella pneumophila, Streptococcus pneumoniae, Haemophilus Influenzae, Mycobacterium tuberculosis, Candida albicans,Streptococcus pyogenes,Streptococcus dysgalactiae subspecies equisimilis. Pneumocystis jirovecii (PJP), Chlamydia pneumoniae, Mycobacterium tuberculosis, Streptococcus salivarius, Pseudomonas aeruginosa, Parainfluenza virus Type2, Parainfluenza virus Type3, Parainfluenza virus Type4, Rhinovirus Human Metapneumovirus (hMPV).

• **Interfering Substances Effect**
A study was performed to evaluate and demonstrate that the endogenous substances naturally present or drugs that may be artificially introduced into clinical samples do not interfere with the detection of SARS-CoV-2, Influenza A, Influenza B, ADV, RSV in the Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit at the concentrations listed below.
Endogenous Substance: Mucin, whole blood, icteric (bilirubin) rheumatoid factor, triglycerides, hemoglobin, anti-nuclear antibody, total IgG, total IgM, total IgA.
Exogenous substance: Mupirocin,Tamiflu (Oseltamivir Phosphate), Fluticasone, Propionate, Fluconazole, Zincum gluconium (i.e., Zicam), Alkalol Phenol, Phenylephrine hydrochloride, Oxymetazolin hydrochloride, Cromolyn Oxymetazoline, Galphimia glauca, Sabadilla Albuterol, Acarbose, Oseltamivir, Chlorpheniramine, Diphenhydramine, Glimepiride (Sulfonylureas), Chlorothiazide,Acetylsalicylic acid, Amoxicillin, Ibuprofen, Beclomethasone, Indapamide, Flunisolide, Guaiaacol glyceryl ether, Biotin, Zanamivir, Tobramycin, Sulfur, Ribavirin, Ephedrine, Benzocaine, Menthol, Budesonide, Triamcinolone, Dexamethasone. Sodium chloride with preservatives,Lopinavir, Ritonavir, Chloroquine phosphate, Ivermectin.

Clinical performance

1. SARS-CoV-2 Test
The performance of Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit were evaluated with 1173 anterior nasal swabs and 1173 Throat swabs. Two swabs were collected with the same people, an anterior nasal swab tested directly using Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit and a Throat swab tested by the RT-PCR Test Kit.

SARS-CoV-2/FluA/FluB+ADV/RSV Antigen Combo Rapid Test Kit	RT-PCR		
	SARS-CoV-2 Positive	Negative	Total
SARS-CoV-2 Positive	102	0	102
Negative	5	1066	1071
Total	107	1066	1173
Diagnostic sensitivity: 95.33% (89.43%~98.47%) Diagnostic specificity: 100.00% (99.65%~100.00%) Accuracy: 99.57%(99.01%~99.86%)			
Positive predictive value: 100.00% (96.45%~100.00%) Negative predictive value: 99.53% (98.91%~99.85%) Kappa:0.9737 95%CI: 0.9508~0.9967			

2. Influenza A Test
The performance of Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit were evaluated with 1173 anterior nasal swabs and 1173 Throat swabs. Two swabs were collected with the same people, an anterior nasal swab tested directly using Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit and a Throat swab tested by the RT-PCR Test Kit.

Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit	RT-PCR		
	Influenza A Positive	Negative	Total
Influenza A Positive	86	0	86
Negative	4	1083	1087
Total	90	1083	1173
Diagnostic sensitivity: 95.56% (89.01%~98.78%) Diagnostic specificity: 100.00% (99.66%~100.00%) Accuracy: 99.66%(99.13%~99.91%)			
Positive predictive value: 100.00% (95.80%~100.00%) Negative predictive value: 99.63% (99.06%~99.90%) Kappa:0.9754 95%CI: 0.9514~0.9995			

3. Influenza B Test
The performance of Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit were evaluated with 1173 anterior nasal swabs and 1173 Throat swabs. Two swabs were collected with the same people, an anterior nasal swab tested directly using Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit and a Throat swab tested by the RT-PCR Test Kit.

Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit	RT-PCR		
	Influenza B Positive	Negative	Total
Influenza B Positive	94	0	94
Negative	5	1074	1079
Total	99	1074	1173
Diagnostic sensitivity: 94.95% (88.61%~98.34%) Diagnostic specificity: 100.00% (99.64%~100.00%) Accuracy: 99.57%(99.01%~99.86%)			
Positive predictive value: 100.00% (96.15%~100.00%) Negative predictive value: 99.54% (98.92%~99.85%) Kappa: 0.9718 95%CI: 0.9471~0.9965			

4. ADV Test
The performance of Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit were evaluated with 1173 anterior nasal swabs and 1173 Throat swabs. Two swabs were collected with the same people, an anterior nasal swab tested directly using Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit and a Throat swab tested by the RT-PCR Test Kit.

Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit	RT-PCR		
	ADV Positive	Negative	Total
ADV Positive	135	0	135
Negative	7	1031	1038
Total	142	1031	1173
Diagnostic sensitivity: 95.07% (90.11%~98.00%) Diagnostic specificity: 100% (99.64%~100%) Accuracy: 99.39%(98.77%~99.76%)			
Positive predictive value: 100% (97.30%~100%) Negative predictive value: 99.39% (98.62%~99.73%) Kappa: 0.9713 95%CI: 0.9502~0.9925			

5. RSV Test
The performance of Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit were evaluated with 1173 anterior nasal swabs and 1173 Throat swabs. Two swabs were collected with the same people, an anterior nasal swab tested directly using Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit and a Throat swab tested by the RT-PCR Test Kit.

Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit	RT-PCR		
	RSV Positive	Negative	Total
RSV Positive	210	0	210
Negative	8	955	963
Total	218	955	1173
Diagnostic sensitivity: 96.33% (92.90%~98.40%) Diagnostic specificity: 100% (99.61%~100%) Accuracy: 99.32%(98.66%~99.71%)			
Positive predictive value: 100% (98.26%~100%) Negative predictive value: 99.17% (98.37%~99.64%) Kappa: 0.9771 95%CI: 0.9614~0.9929			

Usability Study
Usability Study has been conducted in Poland.
Covid-19/ RSV /ADV Influenza A & B Antigen Rapid Test Kit was used to compared with marketed CE marked RT-PCR method. 210 lay persons performed the self-testing and interpreted the result ,which 150 (30 per analyte) were enrolled in diagnostic sensitivity study and 60 were enrolled in diagnostic specificity study. The test result showed:Diagnostic specificity of the product was 100%(60/60). Diagnostic sensitivity of the product for per analyte: SARS-CoV-2: 100%(30/30), Flu A: 96.67%(29/30), Flu B: 96.67%(29/30) , RSV: 96.67% (29/30), ADV: 96.67%(29/30).

Warnings and Precautions

- This test kit is used for in vitro diagnosis only.
- This test kit is for self-testing.
- This test kit should be used within 1 hour after opening the package. The test cassette should not be used if being wet or polluted.
- Proper protection should be taken during testing to avoid splashing when adding sample.
- Dispose of all used or damaged test cassettes, sampling tubes, droppers, swabs, or other kit components as biohazardous materials.
- Negative results do not rule out SARS-CoV-2, Influenza A, Influenza B, ADV, RSV infection, particularly in those who have been in contact with the virus.
- Bring the kit contents to room temperature before testing.
- The test is less reliable when used in the condition of later phase of infection. If testing is not performed within the first 7 days symptom onset, false SARS-CoV-2 negative results may occur. If testing is not performed within the first 4 days symptom onset, false influenza negative results may occur.
- Recommend repeat testing (e.g. within 1-3 days) if there is an ongoing suspicion of COVID/ Influenza infection, being in a high risk setting or where there is an occupational risk or other requirement.
- If the test result is positive, there is a suspicion of ADV or RSV infection, and individuals are advised to consult a medical practitioner for follow-up care.
- This test kit can be used independently by individuals who are 18 or older. For those under the age of 18, it should be operated or supervised by an adult.
- A negative result does not mean a person is not infectious or does not have influenza. If symptoms persist the person should seek medical attention and further testing by PCR if required.
- A negative result does not rule out infection with another type of respiratory virus.
- A positive result cannot necessarily determine whether a person is infectious.

Report Performance or Usability Issues: Contact TGA for calling 1800 809 361 to report poor performance or usability issues in the self-test environment. Report an issue via the Users Medical Device Incident Report, email: iris@health.gov.au

References

1. LY Wang, PR Chen, G W Zheng, et al. Research progress on novel coronavirus test methods. Modern Medicine and Clinic, 2020, 35(3): 411-416.
2. K Tugba, W Ralph, L Hakho. Molecular and Immunological Diagnostic Tests of COVID-19: Current Status and Challenges. IScience, 2020, 23 (8): Doi: 10.1016/j.isci.2020.101406
3. WHO recommendations on the use of rapid testing for inflluenza diagnosis, World Health Organisation, July 2005.
- 4.Tian X , Fan Y , Wang C , et al. Seroprevalence of Neutralizing Antibodies against Six Human Adenovirus Types Indicates the Low Level of Herd Immunity in Young Children from Guangzhou, China[J]. Chinese Journal of Virology: English Edition, 2021, 36(3):9.
- 5.Jing J , Chen Y , Wang Z . Specific IgG antibodies against F and G glycoproteins of respiratory syncytial virus (RSV) in asthmatic children after infection with the virus[J]. Chinexe Journal of Pediatrics, 1998.



Jiangsu Medomics Medical Technology Co., Ltd.
Building 01, Phase 6, No.71, Xinghui Road, Jiangbei New Area, Nanjing, 210000, Jiangsu, P.R. China
Tel: (+86)025-58601060/(+86)025-58601213
Fax: 025-58601060
Web: www.medomics-dx.net
E-mail: overseas@medomics-dx.com / support@medomics-dx.com



Australia & New Zealand Health Products Pty Ltd
Add: Level 8/ 167-169, Queen Street, Melbourne VIC 3000
Web: anzhpl.com.au | pureliving.net.au
E-mail: info@anzhpl.com.au
Telephone: +61 3 9640 0566



Pale Blue Medical Trading Pty Ltd
Add: 5/44 Ellingworth Parade Box Hill, Vic 3128
Web: www.palebluemedical.com.au
E-mail: info@palebluemedical.com.au

ReV1.0.01B.AU03
Effective date: 20/04/2026

IVD

In Vitro Diagnostic Medical Device

LOT

Batch code

REF

Catalogue Number

STERILE EO

Sterilized using ethylene oxide

Do Not Re-use

Do not use if package is damaged

Caution

Contains sufficient for <n> tests

Consult instructions for use

