

# COMMONWEALTH OF AUSTRALIA

## ***Therapeutic Goods Act 1989***

## **MEDICAL DEVICE STANDARDS ORDER (STANDARDS FOR BIOLOGICAL SAFETY OF MEDICAL DEVICES) 2008**

I, LARRY KELLY, delegate of the Minister for Health and Ageing for the purposes of section 41CB of the *Therapeutic Goods Act 1989* and acting under that section,  
DETERMINE:

- (a) that the matters in the relevant standards that are specified in column 2 of an item in the Schedule constitute a medical device standard for the biological safety of a kind of medical device, subject to the conditions (if any) set out in column 3 of that item of the Schedule, and
- (b) that medical devices of those kinds that comply with the appropriate standard specified in column 2 are to be treated as complying with those parts of the essential principles set out in Schedule 1 of the Therapeutic Goods (Medical Devices) Regulations 2002 and that are specified in column 4 of the relevant item of the relevant Schedule.

This Order commences on the day after it is registered on the Federal Register of Legislative Instruments.

Dated this \_\_\_\_\_ day of \_\_\_\_\_ 2008

Larry Kelly  
Delegate of the Minister for Health and Ageing

## Schedule

1 Item No.	2 Medical Device Standard	3 Conditions	4 Essential Principle
1	ISO 10993-1: 2003 <i>Biological evaluation of medical devices - Part 1 Evaluation and testing</i>		Schedule 1, paragraph 7.1(b)
2	ISO 10993-3: 2003 <i>Biological evaluation of medical devices - Part 3 Tests for genotoxicity, carcinogenicity and reproductive toxicity</i>		Schedule 1, paragraph 7.1(b)
3	ISO 10993-4: 2002 <i>Biological evaluation of medical devices - Part 4 Selection of tests for interactions with blood</i>  AND  ISO 10993-4: 2002 <i>Biological evaluation of medical devices - Part 4 Selection of tests for interactions with blood</i> Amendment 1:2006		Schedule 1, paragraph 7.1(b)
4	ISO 10993-5: 1999 <i>Biological evaluation of medical devices - Part 5 Tests for in vitro cytotoxicity</i>		Schedule 1, paragraph 7.1(b)
5	ISO 10993-6: 2007 <i>Biological evaluation of medical devices - Part 6 Tests for local effects after implantation</i>		Schedule 1, paragraph 7.1(b)
6	ISO 10993-7: 1995 <i>Biological evaluation of medical devices - Part 7 Ethylene oxide sterilization residuals</i>		Schedule 1, paragraph 7.1(b)
7	ISO 10993-9: 1999 <i>Biological evaluation of medical devices - Part 9 Framework for identification and quantification of potential degradation products</i>		Schedule 1, paragraph 7.1(b)

1 Item No.	2 Medical Device Standard	3 Conditions	4 Essential Principle
8	<p>ISO 10993-10:2002 <i>Biological evaluation of medical devices - Part 10 Tests for irritation and delayed-type hypersensitivity</i></p> <p>AND</p> <p>ISO 10933-10:2002 <i>Biological evaluation of medical devices - Part 10 Tests for irritation and delayed-type hypersensitivity</i> Amendment 2006</p>		Schedule 1, paragraph 7.1(b)
9	ISO 10993-11: 2006 <i>Biological evaluation of medical devices – Part 11 Tests for systemic toxicity</i>		Schedule 1, paragraph 7.1(b)
10	ISO 10993-12: 2007 <i>Biological evaluation of medical devices - Part 12 Sample preparation and reference materials</i>		Schedule 1, paragraph 7.1(b)
11	ISO 10993-13: 1998 <i>Biological evaluation of medical devices - Part 13 Identification and quantification of degradation products from polymeric medical devices</i>		Schedule 1, paragraph 7.1(b)
12	ISO 10993-14: 2001 <i>Biological evaluation of medical devices - Part 14 Identification and quantification of degradation products from ceramics</i>		Schedule 1, paragraph 7.1(b)
13	ISO 10993-15: 2000 <i>Biological evaluation of medical devices - Part 15 Identification and quantification of degradation products from metals and alloys</i>		Schedule 1, paragraph 7.1(b)
14	ISO 10993-16: 1997 <i>Biological evaluation of medical devices - Part 16 Toxicokinetic study design for degradation products and leachables</i>		Schedule 1, paragraph 7.1(b)
15	ISO 10993-17: 2002 <i>Biological evaluation of medical devices – Part 17 Establishment of allowable limits for leachable substances</i>		Schedule 1, paragraph 7.1(b)
16	ISO 10993-18: 2005 <i>Biological evaluation of medical devices – Part 18 Chemical characterization of materials</i>		Schedule 1, paragraph 7.1(b)