IRIS Program DIRE Committee Results	DOCUME
DIR #: 20799 Device Name: Femoral Component	File No: File Location:
Classification: Urgent: Expedite: Routine:	☐ Not Investigated: ☐
Resolution: Investigate Referral to	
Information Only Await sponsor res	sponse and review
Attendees: AECS MDAS LABS	CLINICAL MVMS
Investigator: Competito	r report (do not send letters)
Letters: Yes No No	
Is the sample going to be tested by Lab staff? Yes	s No No
Name of person testing the sample:	
Special Instructions on Letters: (to be completed (Please provide any additional questions for letters and tick in the brackets if you required)	선거님 [6] 이 나이를 하는 때 보니 요즘은 시간 얼마그런데 본 그 사람들은 발생님들이 되었다고 있다.
 Sample of product () – if requesting samples please tick above Product Specifications () Descriptive product promotional documentation () Instructions for use, as supplied with the device () Device Packaging with printed instructions () Operator's manual () Technical Service Manual () Clinical training manual in printed or video form () In-house training documentation () Evidence of compliance with the Essential Principles () A summary of risk management activities performed by the man Management Report required by Clause 8 of ISO 14971:2000 (nufacturer for the device, e.g. Risk
DIDE C	
- being booked at by the orthogonap.	operation expert working

IRIS Program DIRE Committee Results

45 (<u>1</u> 13. 38)	As a collision et le la collection substitute et de l'est l'estimate de l'abbasse a l'accept. T	. Kristi andra	O progression of the Progress of the Control of the	
Cau	ise of Incident	Res	ult of Investigation	
	Biocompatibility		Bulletin Article	
	Component Failure		Company Warned	
	Contamination		Compliance Testing	
	Design		No Further Action	
	Diagnostic Inaccuracy	\square	Not Investigated	
	Electrical		Other	
	Inadequate Instructions		Problem Not Confirmed	
	Labelling		Product Improvement	
	Maintenance		Recall / Hazard Alert	
	Manufacture		Refer to ADRAC	
	Material / Formulation Deficiency		Refer to GMP	
	Mechanical		Refer to Surveillance	
	Not Applicable - ADR		Safety Alert	
	Not Device Related		User Education	
	Other		,	
	Packaging / Sterility			
	Quality Assurance			
	Unknown			
	Wear / Deterioration			
Recommendation (please circle the appropriate recommendation): 1. The cause of this problem has not been conclusively determined; however the low level of occurrence is not				
۱.	currently cause for concern.	CCII	conclusively determined, however the low level of occurrence is not	
2.)	No further investigation will occur		is time, however the TGA will continue to monitor the rate and pattern of	
3.	occurrence and may re-open the file		appropriate. o not indicate further investigation is appropriate at this time. The TGA	
J.			dents and may re-open the file if appropriate.	
4.	This product is not a therapeutic de		within the meaning of the Therapeutic Goods Act 1989. This report will	
5.	not be investigated. The framework for the regulation of	f dis	infectants and sterilants is not fully implemented. Sponsors have until	
•	October 1998 to fully comply with	labe	ling requirement.	
6.	This report is entered for the record, and is not investigated at this time. The incidence of failure with this device			
7	is reviewed regularly for changes in trending or significant failure modes. The information is entered for the record. This report is not being investigated at this time.			

Other recommendations:

Form completed by: