



Australian Government  
Department of Health and Ageing  
Therapeutic Goods Administration

## LIMS Entry Summary

ARTG Number: 114907  
ARTG Name: Poly Implant Protheses - High Cohesive Gel Pre-Filled Breast Implant (Textured)  
IMGHC-TX-S  
Sponsor: Medical Vision Australia Pty Ltd  
Batch Number: 25109

The sample was submitted by Sponsor

Results Approved By: Jorge Garcia  
Results of the examination and analysis were as follows:

**Approval Comments:**

The sample is NOT cytotoxic

TEST	RESULT	COMMENTS	ENTERED BY	VALIDATED BY
<b>CYTOTOXICITY med dev</b>				
- S1	>100		wirasr	garcij
- S2	100		wirasr	garcij
- S3	>100		wirasr	garcij
- Positive control	13		wirasr	garcij
- Negative control	>100		wirasr	garcij
- Blank control	>100		wirasr	garcij
- Cytotoxic titre	100		wirasr	garcij