



Australian Government
Department of Health and Ageing
Therapeutic Goods Administration

LIMS Entry Summary

ARTG Number: 114908
ARTG Name: Poly Implant Protheses- High Cohesive Gel Pre-Filled Breast Implant (Smooth Texture)
IMGHC-LS-S
Sponsor: Medical Vision Australia Pty Ltd
Batch Number: 56206

The sample was submitted by Sponsor

Results Approved By: Jorge Garcia Sample Status: VA

Results of the examination and analysis were as follows:

Approval Comments:

The Sample is NOT cytotoxic

TEST	RESULT	COMMENTS	ENTERED BY	VALIDATED BY
CYTOTOXICITY med dev				
- S1	>100		carsca	garcij
- S2	>100		carsca	garcij
- S3	>100		carsca	garcij
- Positive control	20		carsca	garcij
- Negative control	>100		carsca	garcij
- Blank control	>100		carsca	garcij
- Cytotoxic titre	>100		carsca	garcij