



Operations Biotherapeutics Laboratories Operations Manual

Procedure Identification of the mRNA in modRNA BNT162b2 (1525) using RT-PCR Assay

Written [REDACTED]

Authorised [REDACTED]

Date issued 03 February 2021

Revision # 1

TRIM link to SOP	D21-2063055	Date of assay	19/07/2021
TRIM link to PCR template	D21-2061428	TRIM link to assay data file	el://D21-2860646?db=A7&open
Operator	[REDACTED]	Checked by	[REDACTED]

Reaction mixture preparation: Prepare 15µL of reaction mixture per well plus 10% overage

Primers and probe are provided by Pfizer and are stored in aliquots at working concentration (x10)

Each assay requires a total of 12 wells for PCR and extraction controls plus 3 wells per test sample.

Reporting: Record the sample LIMS numbers in the table below prior to the assay. Once the assay and analysis are complete, record the assay validity criteria parameters on the worksheet below noting whether the assay is valid. For a valid assay, record the sample Ct values and whether identity has been confirmed in the table below.

Using the Quantstudio Design & Analysis Software, update the assay results file to include the LIMS numbers of each sample (replacing the placeholder letter designation), using the "Advanced Setup" pane of the "Plate" tab. Save the updated file in the assay specific data folder in TRIM (E21-219384).

Following completion of the assay, convert this worksheet into a .pdf and append to it the following: .pdf copies of amplification curve plots, a .pdf experiment report as generated by the Quanstudio Design & Analysis Software. Combine these into a single .pdf and file it in the assay specific data folder (E21-219384) along with the Quantstudio data file.

Reagent information

Reagent	Manufacturer and Catalogue Number	Lot Number	Expiry	Notes
QIAamp Viral RNA Mini Kit	Qiagen, 52906	166024851	2022-03-09	
Buffer AVL	Qiagen, 52906	166024291	2022-04-21	
Buffers AW1 and AW2	Qiagen, 52906	166023682 & 166019545	N/A	
Ethanol 96-100%	Supelco, 1.00983.2511	K52687883 035	2025/06/30	
TE Buffer	Life Technologies, AM9849	01015399	N/A	
RT-PCR-grade water	Life Technologies, AM9935	2004108	N/A	
TaqPath 1-step RT-qPCR Master Mix, CG	Applied BioSystems, A15299	2255145	N/A	
ModRNA1525 RT-PCR forward primer working stock	Pfizer, 4304972	00711497-0162-M06		see last page for probe and primer details
ModRNA1525 RT-PCR reverse primer working stock	Pfizer, 4304972	00711497-0162-M04		
ModRNA1525 RT-PCR probe working stock	Pfizer, 4316032	00711497-0162-M02		

Record Details Identification of the mRNA in modRNA BNT162b2 (1525) using RT-PCR Assay

Last Editor [REDACTED] Edit Date 20/07/2021 8:44 AM

Print Date 20/07/2021 8:44 AM

Page 1 of 4

Sample Preparation and Dilution

Positive PCR Control	Positive Extraction Control	Negative Extraction Control	Test Samples
Dilution 1: 10µL DS RM (10µg/mL) + 90µL RT-PCR-grade water = 1µg/mL Dilution 2: 10µL Dilution 1 (1µg/mL) + 190µL RT-PCR-grade water = 50ng/mL = 50 pg/µL 5µL Dilution 2 per RT-PCR well	Prepare with QIAamp Viral RNA Mini Kit. Lysis step composition: 140µL undiluted DP RM + 560 QIAamp Buffer AVL, without carrier RNA. Follow kit protocol Elute in 60µL Buffer Dilution 1: 10µL eluate + 990µL RT-PCR-grade water Dilution 2: 10µL Dilution 1 + 990µL RT-PCR-grade water 5µL Dilution 2 per RT-PCR well	Prepare with QIAamp Viral RNA Mini Kit. Lysis step composition: 140µL TE buffer + 560 QIAamp Buffer AVL, without carrier RNA Follow kit protocol Elute in 60µL Buffer Dilution 1: 10µL eluate + 990µL RT-PCR-grade water Dilution 2: 10µL Dilution 1 + 990µL RT-PCR-grade water 5µL Dilution 2 per RT-PCR well	Prepare with QIAamp Viral RNA Mini Kit. Lysis step composition: 140µL undiluted test sample + 560 QIAamp Buffer AVL, without carrier RNA Follow kit protocol Elute in 60µL Buffer Dilution 1: 10µL eluate + 990µL RT-PCR-grade water Dilution 2: 10µL Dilution 1 + 990µL RT-PCR-grade water 5µL Dilution 2 per RT-PCR well

RT-PCR Reaction Mixture

Number of wells required	15	Volume of master mix per well	15µL	Total volume of master mix required (incl. 10% overage)	255	Volume of sample per well	5µL
--------------------------	----	-------------------------------	------	---	-----	---------------------------	-----

Component	Volume per well	Total Volume (incl. 10% overage)
Mastermix	10 µL	170
Forward Primer	1 µL	17
Reverse Primer	1 µL	17
Probe	1 µL	17
RT-PCR grade water	2 µL	34
Total volume of reaction mixture	255	

Reaction mixture notes:

Mastermix used: Taqman Fast Advanced / Taqman Universal II

4X TaqPath 1-step Mastermix, CG

Record Details	Identification of the mRNA in modRNA BNT162b2 (1525) using RT-PCR Assay		
Last Editor		Edit Date	20/07/2021 8:44 AM
Print Date	20/07/2021 8:44 AM		Page 2 of 4

PCR Plate layout:

	1	2	3	4	5	6	7	8	9	10	11	12
A	PCR Positive Control	PCR Positive Control	PCR Positive Control							Positive Extraction Control	Positive Extraction Control	Positive Extraction Control
B												
C	Sample A	Sample A	Sample A									
D	Sample B	Sample B	Sample B									
E	Sample C	Sample C	Sample C									
F	Sample D	Sample D	Sample D									
G												
H	No Template Control	No Template Control	No Template Control							Negative Extraction Control	Negative Extraction Control	Negative Extraction Control

Sample	Sample LIMS#:	Ct Value	Identity results (confirmed/not confirmed) (Ct must be < 32 for all sample replicates)
Sample A	2107002694	23.186, 21.432, 21.128	confirmed
Sample B			
Sample C			
Sample D			

Record Details	Identification of the mRNA in modRNA BNT162b2 (1525) using RT-PCR Assay		
Last Editor		Edit Date	20/07/2021 8:44 AM
Print Date	20/07/2021 8:44 AM		Page 3 of 4

Assay Validity Data

Criterion	Required value	Observed values	Validity
Ct of each Negative Extraction Control well	> 32 or undetermined	All undetermined	Valid
Ct of each No Template Control well	> 32 or undetermined	All undetermined	Valid
Ct of each PCR Positive Control well	< 32	17.479, 17.076, 16.557	Valid
Ct of each Positive Extraction Control well	< 32	17.546, 17.453, 17.770	Valid

Sample interpretation

Sequence identity of the mRNA is considered confirmed if all test sample replicates show amplification curves with a Ct value of < 32.0000. If all test sample replicates show amplification curves with a Ct value of > 32.0000 the identity is considered not confirmed. If a mixture of results (with Ct values both greater and less than 32.0000) is found for a test sample it must be repeated.

Pipettes used:

33166, 33190, 30006, 5646, 5653, 33087

Probe: Applied bioSystems, supplied by ThermoFisher, Lot# 7495294-1 C1, Ref# 4316034

Forward Primer: Merck, supplied by Sigma, Ref# VC00021, SY21020241530-088, Lot# 3026595983-000020


Reverse Primer: Merck, supplied by Sigma, Ref# VC00021, SY21020241529-078, Lot# 3026595983-000030

100uM Probe stock prepared as per manufacturer's instructions, 10uL of 100uM added to 190uL RT-PCR Grade water to make 5uM stock, MJ 19Jul21

100uM Primer stock prepared as per manufacturer's instructions, 20uL of 100uM added to 91uL RT-PCR Grade water to make 18 uM stock

Record Details	Identification of the mRNA in modRNA BNT162b2 (1525) using RT-PCR Assay		
Last Editor		Edit Date	20/07/2021 8:44 AM
Print Date	20/07/2021 8:44 AM		Page 4 of 4

Amplification Plot - BNT162b2 Identification assay - 18Jul21



Experiment Summary

Experiment Name: 2021-01-11_133749
Experiment Type: Standard Curve
Chemistry: TaqMan® Reagents
BarCode:
File Name: ID Pfizer.eds
Run Started: 07-19-2021 17:54:35 AEST
Run Finished: 07-19-2021 19:00:44 AEST
Run Duration: 66 minutes 8 seconds
Date EDS Modified: 07-20-2021 07:15:42 AEST
Date EDS Created: 01-11-2021 15:21:26 AEDT
User Name:
Number of Wells Used: 15
Number of Wells with Results: 15
Instrument Name: QS3
Instrument Type: QuantStudio™ 3 System
Instrument Serial Number: 272322852
Model/Block Type: QuantStudio™ 3 System / 96-Well 0.2-mL Block
Quantification Cycle Setting: CT
Stage/Step for Analysis: Stage2, Step2
Comments:



Printed:

07-20-2021 08:57:28 AEST

Reagent Information

Type	Name	Part Number	Lot Number	Expiration Date



Printed: 07-20-2021 08:57:28 AEST

2 / 6

Results Summary

Sample	Target	Qty Mean	Qty SD	Ct Mean	Ct SD
2107002694	Target 1	0.000		21.915	1.111
NEC	Target 1				
NTC	Target 1				
PEC	Target 1	1.000		17.590	0.163
Positive Control	Target 1	1.000		17.038	0.462



Printed: 07-20-2021 08:57:28 AEST

3 / 6

For Research Use Only, not for use in diagnostic procedures.

Experiment Results Report

Plate Layout

	1	2	3	4	5	6	7	8	9	10	11	12
A	Positive.. Target 1	Positive.. Target 1	Positive.. Target 1							PEC Target 1	PEC Target 1	PEC Target 1
B												
C	1 210700266 ----- T-----A	1 210700266 ----- T-----A	1 210700266 ----- T-----A									
D												
E												
F												
G												
H	NEC Target 1	NEC Target 1	NEC Target 1						NTC Target 1	NTC Target 1	NTC Target 1	

RUO

For Research Use Only, not for use in diagnostic procedures.

Printed:

07-20-2021 08:57:28 AEST

Experiment Results Report

Results Table

Well	Sample	Target	Task	C _T	C _T Mean	C _T SD	Qty	Qty Mean	Qty SD	C _T Threshold	Baseline Start	Baseline End	C _q Conf
A1	Positive Control	Target 1	S	17.479	17.038	0.462	1.000	1.000		0.087	3	12	0.928
A2	Positive Control	Target 1	S	17.076	17.038	0.462	1.000	1.000		0.087	3	12	0.942
A3	Positive Control	Target 1	S	16.557	17.038	0.462	1.000	1.000		0.087	3	12	0.936
H11	NTC	Target 1	N	UND.						0.087	3	39	0.000
H10	NTC	Target 1	N	UND.						0.087	3	39	0.000
A10	PEC	Target 1	S	17.546	17.590	0.163	1.000	1.000		0.087	3	13	0.959
A11	PEC	Target 1	S	17.453	17.590	0.163	1.000	1.000		0.087	3	12	0.955
A12	PEC	Target 1	S	17.770	17.590	0.163	1.000	1.000		0.087	3	13	0.952
H12	NTC	Target 1	N	UND.						0.087	3	39	0.000
C1	2107002694	Target 1	U	23.186	21.915	1.111	0.000	0.000		0.087	3	18	0.950
C2	2107002694	Target 1	U	21.432	21.915	1.111	0.000	0.000		0.087	3	16	0.959
C3	2107002694	Target 1	U	21.128	21.915	1.111	0.000	0.000		0.087	3	16	0.952
H1	NEC	Target 1	N	UND.						0.087	3	39	0.000
H3	NEC	Target 1	N	UND.						0.087	3	39	0.000
H2	NEC	Target 1	N	UND.						0.087	3	39	0.000

Legend: S = Standard, N = Non Template Control, U = Unknown, UND. = Undetermined



For Research Use Only, not for use in diagnostic procedures.

Printed:

07-20-2021 08:57:28 AEST

QC Summary

Total Wells:96 Processed Wells:15 Targets Used:1 Well Setup:15 Flagged Wells:3 Samples Used:5

Flag	Description	Frequency	Locations
AMPNC	Amplification in negative control	0	
BADROX	Bad passive reference signal	0	
BLFAIL	Baseline algorithm failed	0	
CQCONF	Low Cq confidence	0	
CTFAIL	Ct algorithm failed	0	
DRNMIN	Define acceptable delta Rn based on Ct range	0	
EXPFAIL	Exponential algorithm failed	0	
HIGHSD	High standard deviation in replicate group	3	C1, C2, C3
HIGHSD	High standard deviation in replicate group	0	
NOAMP	No amplification	0	
NOISE	Noise higher than others in plate	0	
NOSIGNAL	No signal in well	0	
OFFSCALE	Fluorescence is offscale	0	
OUTLIERRG	Outlier in replicate group	0	
PRFDROP	Passive reference signal changes near Ct	0	
PRFLOW	Low passive reference signal	0	
SPIKE	Noise spikes	0	
THOLDFAIL	Thresholding algorithm failed	0	

Customer is responsible for any validation of assays and compliance with regulatory requirements that pertain to their procedures and uses of reagents and/or instruments. ©2018 Thermo Fisher Scientific. All rights reserved. The trademarks mentioned herein are the property of Thermo Fisher Scientific or their respective owners.



Printed:

07-20-2021 08:57:28 AEST

6 / 6