



IMDRF International Medical
Device Regulators Forum

FINAL DRAFT
DOCUMENT

International Medical Device Regulators Forum

Title: Assembly and Technical Guide for IMDRF Table of Contents (ToC)
Submissions (ToC-based submissions)

Authoring Group: Regulated Product Submissions Table of Contents WG

Date: September 17, 2015

Table of Contents

1	1.0	Introduction.....	4
2	2.0	Scope.....	4
3	3.0	GUIDE TO BUILDING A TOC-BASED SUBMISSION.....	5
4	3.1	Pilot Documents.....	5
5	3.2	Sample general process for building a ToC-based submission.....	6
6	4.0	TECHNICAL GUIDELINES.....	8
7	4.1	Folder Structure.....	8
8	4.2	Folder Naming Convention.....	9
9	4.3	File Format and Naming.....	10
10	4.4	File and Submission Size Limitations.....	11
11	4.5	Document Security.....	11
12	4.6	Bookmarking in PDF Files.....	11
13	4.7	Hyperlinking in PDF files.....	11
14	4.8	Granularity Rules.....	12
15	4.9	Pagination.....	12

16 **Preface**

17

18 The document herein was produced by the International Medical Device Regulators Forum
19 (IMDRF), a voluntary group of medical device regulators from around the world. The document
20 has been subject to consultation throughout its development.

21

22 There are no restrictions on the reproduction, distribution or use of this document; however,
23 incorporation of this document, in part or in whole, into any other document, or its translation
24 into languages other than English, does not convey or represent an endorsement of any kind by
25 the International Medical Device Regulators Forum.

26 **1.0 Introduction**

27 The International Medical Device Regulators Forum (IMDRF) was conceived in February 2011
28 as a forum to discuss future directions in medical device regulatory harmonization. It is a
29 voluntary group of medical device regulators from around the world who have come together to
30 build on the strong foundational work of the Global Harmonization Task Force (GHTF). The
31 Forum will accelerate international medical device regulatory harmonization and convergence.
32 The Regulated Product Submission (RPS) proposal was endorsed as a New Work Item (NWI) by
33 IMDRF at its inaugural meeting in Singapore (March 2012). The working group to this point
34 has accomplished the following:

- 35 1. Established that the Health Level Seven (HL7) RPS Standard is "fit for purpose" for the
36 electronic exchange of information related to premarket medical device applications.
- 37 2. **Established a comprehensive Table of Contents (ToC) for the following premarket**
38 **applications**
 - 39 a. **Non-IVD (nIVD) Market Authorization**
 - 40 b. **IVD Market Authorization**

41 The ToC Working Group¹ has previously conducted pilots for both the nIVD and IVD Market
42 Authorization ToC structures, using historical submissions. These pilots provided valuable
43 feedback regarding the ToC structure and completeness, however there were obvious limitations
44 to using historical submissions and there were limited samples involving more than one region.
45 Furthermore, there were no specific guidelines regarding the means of building a submission in a
46 pre-RPS implementation.

47 This document is intended to supplement the IMDRF ToC Pilot Plan and describe additional
48 harmonized guidelines for the acceptable folder structure and file format(s) for ToC-based
49 submissions.

50 **2.0 Scope**

51 This guide is intended for use in the assembly of IMDRF Table of Contents (ToC) based medical
52 device regulatory submissions currently within the scope of submission types accepted by each
53 IMDRF region.

¹ The IMDRF Table of Content Working Group is composed of the regulatory authorities from the agencies represented by the IMDRF Management Committee.

54 **3.0 GUIDE TO BUILDING A TOC-BASED SUBMISSION**

55 There are a number of reference documents and guides that need to be consulted when creating a
 56 ToC-based medical device submission. This section provides information about these reference
 57 documents as well as information about how to use these documents to generate a ToC-based
 58 submission.

59 **3.1 Pilot Documents**

60 The table below lists the documents required to assemble an IMDRF ToC-based regulatory
 61 submission during the IMDRF TOC Pilot.

62 Table 1 - List of pilot documents

Document	Description	Location
IMDRF In Vitro Diagnostic Medical Device Market Authorization Table of Contents (IVD MA ToC) or IMDRF Non-In Vitro Diagnostic Medical Device Market Authorization Table of Contents (nIVD MA ToC)	These documents define the heading names and hierarchy of the ToC structure. They also include detailed information about the content that belongs under each heading.	www.imdrf.org
IMDRF Assembly and Technical Guide for IMDRF Table of Content (ToC) Submissions [THIS DOCUMENT]	This document provides information about the reference documents available relating to the IMDRF ToC and harmonized technical specifications for ToC-based submissions.	www.imdrf.org (when finalized)
IMDRF Frequently asked Questions Document	Additional reference document that provides responses to commonly asked questions.	www.imdrf.org (when available)
IMDRF Standard ToC Folder Structure (presented as a zip file)	This is a folder structure provided by IMDRF to replicate the hierarchy and headings of the ToC. Note: some headings have been modified from the full names defined in the nIVD and IVD MA ToC documents to reduce path lengths.	www.imdrf.org

Document	Description	Location
<u>REGIONAL</u> Classification Matrix	As the IMDRF ToC documents are comprehensive in nature, not all headings are required for all submission types and/or regions. The classification matrix defines whether for the given submissions type a heading is required, not required, optional, conditionally required, etc.	Various - contact region* of interest for details
<u>REGIONAL</u> Assembly and Technical Guide for IMDRF Table of Content (ToC) Submissions	Similar to this document, regions may have additional requirements or regional specific guidance relating to the building and submission of a ToC-based submission that will be included in a regional Assembly and Technical Guide (e.g. transmission methods or special instructions for file transfer media).	Various - contact region* of interest for details
<u>REGIONAL</u> Frequently asked Questions Document	Additional reference document that provides responses to commonly asked questions.	Various - contact region* of interest for details

*The regions involved in the pilot are ANVISA, EU, Health Canada, TGA, USFDA and CFDA.

63 3.2 Sample general process for building a ToC-based submission

64 This section describes one example of how the pilot documents could be used to manually
65 compile an IMDRF ToC pilot submission. It is important to note that this is intended to provide
66 further context to the pilot documents. Other approaches may be acceptable, including using
67 commercially available submission publishing software to generate a submission meeting the
68 requirements defined in the pilot documents.

69 **Step 1:** Download² the required IMDRF Standard ToC Folder Structure for the applicable ToC
70 structure (e.g. IVD or nIVD)

71 **Step 2:**

72 · **Step 2a:** Begin building the submission consulting the relevant IMDRF Market
73 Authorization Table of Contents (IVD MA ToC or nIVD MA ToC) for content related
74 guidance. Consult the regional classification matrix to establish the headings that require
75 content based on the submission type. **See Implementation Considerations below for**
76 **important considerations in this process.**
77

² See IMDRF Standard ToC Folder Structure file

78 · **Step 2b:** Consult this document as well as the IMDRF FAQ documents and regional
79 equivalents for the region of interest for technical requirements relating to the
80 files/folders that must be populated.

81 **Step 3:** Consult the regional classification matrix of interest to establish which folders can be
82 deleted from the comprehensive structure based on the submission type – see **Section 4.1 below**
83 **for further guidance.**

84 Implementation Considerations:

- 85 1. Implementers should consider the potential for maintaining content that will be submitted
86 to multiple regions. Although certain regions may have additional content requirements
87 for certain headings³, it may be prudent to build a non region-specific version of the
88 submission using the complete IMDRF Standard ToC Folder Structure⁴. Make copies of
89 this complete version for each region that you are intending to submit before deleting any
90 folders not required for the intended region. Future regional adaptations can then be more
91 easily produced from this baseline submission structure and content. This reduces the risk
92 of:
- 93 ○ Inclusion of regional content that is not required for the submission.
 - 94 ○ Missing required elements due to folders that were deleted but are required for
95 other regions.
- 96 2. Conversely, if the approach described above is not possible and a submission is being
97 built from a folder structure previously submitted to another region, take care to:
- 98 ○ Consider those heading that are regional or require regional focus and to ensure
99 that regional content that is not relevant to the subject regulator is removed.
 - 100 ○ Ensure that any folders that may have been deleted for the original submission are
101 reconsidered for inclusion in the new submission.
 - 102 ○ Ensure that content is current (e.g. market history is up to date).
- 103
- 104

³ For a complete description of common and regional content requirements for each heading refer to: [IMDRF In Vitro Diagnostic Medical Device Market Authorization Table of Contents \(IVD MA ToC\)](#) or [IMDRF Non-In Vitro Diagnostic Device Market Authorization Table of Contents \(nIVD MA ToC\)](#)

⁴ Note that the use of PDFs and folders is an interim format designed to test a standardised Table of Contents and future implementations may be more user friendly. It is time consuming for multiple reviewers to open chains of folders every time to find that they are empty and to ensure there is no evaluable content. It is not normal practice for regulators to alter submissions, e.g. by deleting unneeded documents or folders, as the regulators need to maintain the integrity of the original submission. The IMDRF ToC Working Group appreciates your efforts to delete empty folders and provide content as per the regional guidance documents before submission.

105 4.0 TECHNICAL GUIDELINES

106 The IMDRF ToC Pilot will rely on technical guidelines to provide consistency across the
107 regions. The following sections include basic guidelines for submitting a ToC based submission.

108 4.1 Folder Structure

109 The IMDRF documents, In Vitro Diagnostic Medical Device Market Authorization Table of
110 Contents (IVD MA ToC) and Non-In Vitro Diagnostic Device Market Authorization Table of
111 Contents (nIVD MA ToC) define the content for each folder. The folder structure is to be built as
112 prescribed by IMDRF. Refer to the IMDRF Standard ToC Folder Structure file, which is a
113 physical folder structure template provided by IMDRF to help facilitate the preparation of
114 applications in the required ToC format.

115 Regional Classification Matrices describe which elements of the ToC are required for each
116 regulatory submission within scope. There are factors influencing the inclusion/exclusion of
117 submission contents, these considerations are detailed below.

118 Each folder within the submission has been established as either **Required** or **Not Required** for
119 the particular submission. This is explicitly defined by the classification matrix (e.g. *Required* or
120 *Not Required* classification) or through interpretation of the classification (e.g. through
121 assessment of conditions⁵ for those that are classified as *Conditionally Required* or a decision by
122 the applicant for those that are classified as *Optional*). With this in mind, Figure 1 below depicts
123 many of the classifications that can result in a folder being **Required** or **Not Required** within
124 the submission.

125 Any folder that is established as **Required** should not be deleted. Content must be submitted in
126 this folder.


127 Any folder that is established as **Not Required** should be deleted to ensure the submission
128 content package does not contain empty folders. If any parent folder contains no content, then
129 that parent folder should also be deleted.

130 Any folder that is established as **Conditionally Required** requires a determination against the
131 conditions by the applicant. A folder should be retained if content is required or should be
132 deleted if content is not required.

133
134 Any folder that is established as **Optional** requires a decision by the applicant then should be
135 deleted if not populated.

136 It should be noted that some regions may require a statement describing why a section is not
137 provided. Refer to regional pilot instructions for additional details.

⁵ Conditions for *Conditionally Required* headings are outlined in the Classification Matrices



138 Figure 1 - Those classifications defined in the classification matrix (rectangles) that can lead to content being
 139 Required or Not Required in for a particular submission (ovals).

140 4.2 Folder Naming Convention

141 The folders in the provided templates will be numbered and named per the ToC requirements,
 142 with the exception of the custom headings which are to be numbered and named as defined in the
 143 IMDRF ToC (e.g. [Study description, study identifier, date of initiation] indicating the
 144 information the sender should include as the folder name of the custom folder. In the IMDRF
 145 Standard Folder Structure file – these folders appear with “[Custom]” in the folder name). The
 146 final digit of the heading number should be revised as appropriate to ensure appropriate
 147 sequential presentation of the custom folders when more than one study is being included.
 148 For example, for the Physical and Mechanical Characterization heading, the first custom study
 149 folder should be named “3.5.01.1[Study description, study identifier, date of initiation]” and the
 150 second custom study folder should be named “3.5.01.2[Study description, study identifier, date
 151 of initiation]”. If there are more than 10 studies, the sequence numbering should use 2 digits
 152 (e.g. 3.5.01.01... 3.5.01.10 for the example above).

153
 154 The character count for the [Custom] or [Trial Details] folder names should be no more than 50
 155 characters (including the section number). Abbreviations in their names are expected and
 156 acceptable

157
 158 **NOTE:** Restrictions in file and folder names exist to ensure maximum allowable system file path
 159 lengths are not exceeded.

160 4.3 File Format and Naming

161 Portable document format (PDF) files are the preferred file format although other formats such
162 as Microsoft Office (.doc, .ppt, .xls) are also acceptable in some regions. Refer to regional pilot
163 instructions for specific file format and versions.

164 The applicant should create all PDF files directly from the source documents whenever feasible
165 rather than creating them by scanning. **PDF documents produced by scanning paper**
166 **documents are far inferior to those produced directly from the source document, such as**
167 **Word document, and, thus, should be avoided if at all possible.** Scanned documents,
168 particularly tables and graphs, are more difficult to read and do not allow the reviewers to copy
169 and paste text.

170 For any scanned document, you should perform optical character recognition (OCR) so that the
171 text is searchable. Check to see that the content has been correctly converted by: (1) highlighting
172 an area of text and (2) searching for a word or phrase. If the word or phrase is not returned in the
173 search, then the OCR did not recognize the text. We recognize that OCR may not be feasible in
174 some cases for documents with figures and images.

175 Most file names are user defined, with a limitation of 50 characters (including the file extension
176 and section number). File names should be meaningful and provide some indication of their
177 content. When more than one file is presented in a folder, suffix number should be used to ensure
178 the intended sequence of presentation is maintained.

179 File names are prescribed under the user defined [Custom] folders and must be:

- 180 • 1-Summary
- 181 • 2-Full Report
- 182 • 3-Statistical Data

183
184 File names are prescribed under the user defined [Trial details] folder and must be:

185

nIVD	IVD
1-Clinical Trial Synopsis	1-Clinical Study Synopsis
2-Clinical Trial Report	2-Clinical Study Report
3-Clinical Trial Data	3-Clinical Study Data

186

187 **NOTE:** Restrictions in file and folder naming exist to ensure maximum allowable system file
188 path lengths are not exceeded. Applicants should be aware that computer operating systems have
189 limitations and are requested to **keep filenames and pathnames in submissions as short as**
190 **possible.**

191
192 The IMDRF folder templates and file naming specifications have been established in an effort to
193 ensure submissions can be received and stored by regulators without reaching the operating
194 system limits. It is recommended that applicants examine the length of the entire pathname (i.e.
195 all nested folders and file name and file extension) prior to transmission to verify the path length
196 is 200 characters or less.

197 **4.4 File and Submission Size Limitations**

198 No individual PDF file in the submission shall exceed 100 MB.

199 The entire submission should not exceed 4GB to ensure acceptance by all participating regions.

200 **4.5 Document Security**

201 Files should not have any security settings, specifically:

202

- 203 • Files must not have password protection preventing the file from opening.
- 204 • Files should be set to allow printing, selecting text and graphics, and adding or changing
- 205 notes and form fields.
- 206

Applicants should use secure upload facilities or reputable couriers to protect the transmission to the regulators.

207 **4.6 Bookmarking in PDF Files**

208 It is also important that PDF files be properly structured, with a properly bookmarked internal
209 table of contents. The following are recommended as good structuring practices:

- 210 • Documents of ten pages or more should have their own internal table of contents.
- 211 • When creating bookmarks, the magnification setting should be set to Inherit Zoom so that
- 212 the destination page displays at the same magnification level that the reviewer is using for
- 213 the rest of the document.
- 214 • Sections, subsections, tables, figures and appendices should all be bookmarked.
- 215 • Attachments to PDF files should be avoided.
- 216 • Too many levels of bookmarks are inefficient. In most instances, three levels of
- 217 bookmarks should be sufficient:
 - 218 1 Heading
 - 219 1.1 Subheading
 - 220 1.1.1 Sub-subheading.

221 It is recognized that bookmarks are generated automatically from document headings;
222 nevertheless, it is recommended that they be kept concise.

223 Set the Navigation Tab to open to “Bookmarks Panel and Page.” This sets the initial document
224 view when the file is opened. If there are no bookmarks, set the Navigation Tab to “Page Only.”
225 Page Layout and Magnification should be set to “Default.”

226 **4.7 Hyperlinking in PDF files**

227 Hyperlinks are used to improve navigation through individual PDF documents and are
228 encouraged. Hyperlinks can be designated by rectangles using thin lines or by blue text or you
229 can use invisible rectangles for hypertext links in a table of contents to avoid obscuring text.
230 Hyperlinks throughout the body of the document to supporting annotations, related sections,

231 references, appendices, tables, or figures that are not located on the same page are helpful and
232 improve navigation efficiency.

233 Hyperlinks between documents are acceptable but care must be taken in creating the links
234 between different documents so that they will function once the application is received by the
235 regulator (the use of relative linking is recommended). It is the applicant's responsibility to
236 ensure that hyperlinks are functioning. Links must also include references to the specific section
237 or page in the event the link is broken.

238 **4.8 Granularity Rules**

239 There are no limitations on the number of files per heading within the submission, however, the
240 following guidelines should be considered.

241 1) Efforts should be made to draft documents that concisely communicate the content
242 described in the IMDRF [In Vitro Diagnostic Medical Device Market Authorization Table](#)
243 [of Contents \(IVD MA ToC\)](#) or [Non-In Vitro Diagnostic Medical Device Market](#)
244 [Authorization Table of Contents \(nlVD MA ToC\)](#), rather than simply including existing
245 documentation that contains superfluous information not required for the particular
246 heading. For example:

247 • Including a number of Material Safety Data Sheets within “*2.4.1 - Comprehensive*
248 *Device Description and Principle of Operation*” is less helpful than summarizing
249 the specific details of relevance to this heading.

250 2) When multiple files are considered necessary, file naming methods should ensure that the
251 files are presented in their intended sequence, for example in folder named “*2.4.1-*
252 *Comprehensive Device Description & Principle of Operation*” the files would appear as:
253 2.4.1.0-Comprehensive Device Description and Principle of Operation.pdf
254 2.4.1.1-Engineering drawings.pdf

255 **4.9 Pagination**

256 Pages of the submission should be numbered in such a manner that information can be easily
257 referenced by page number. Pagination should be applied to each document (i.e., the physical
258 file). This may be done by numbering the pages within a section or chapter (e.g., CH2.4.1-1,
259 CH2.4.1-2).