

Medical Device Incident Investigations: Recommendations

Reprocessing phaco-emulsifying handpieces

The TGA Incident Report Investigation Scheme (IRIS) has received several reports in recent times about concerns with cleaning and sterilising reusable phacoemulsion handpieces used in eye surgeries.

Concerns from theatre staff and sterilising managers relate more specifically to the perceived lack of manufacturers' instructions on how to reprocess these devices if used on patients suspected of having Creutzfeldt-Jakob Disease (CJD) or when used in high risk infectivity sites such as specific parts of the eye. (See Department of Health and Ageing 2004, *Infection control guidelines for the prevention of transmission of infectious diseases in the health care setting* <<http://www.health.gov.au/internet/main/publishing.nsf/Content/icg-guidelines-index.htm>>).

Recommendation

The TGA recommends that if hospital staff suspect that a patient may be infected with a prion-type disease, those responsible for ensuring the safe reprocessing of such devices should firstly contact the sponsor of the device to obtain appropriate manufacturer's reprocessing instructions.

If these instructions are not sufficient to meet the requirements of the healthcare facility's policy and procedures, it is recommended that the healthcare facility then follow local or national guidelines on how to manage these instruments.

MEDICAL DEVICE INCIDENT REPORT INVESTIGATION SCHEME (IRIS) STATISTICS REPORT 01/07/2008 to 30/09/2008

Total Number Received: 397

Cause of Problem ¹	Effect	Result of Investigation
Component Failure	Death	Bulletin Article
Contamination	Serious Injury	Company Warned
Design	Temporary Injury	No Further Action
Electrical	No Injury	Not Investigated ³
Inadequate Instructions		Other
Labelling	Source Category	Problem Not Confirmed
Maintenance	Medical Administrator	Product Improvement
Manufacture	Specialist	Recall/Hazard Alert
Material/Formulation Deficiency	General Practitioner	Safety Alert
Mechanical	Nurse	User Education
Not Device Related	Blood Bank	
Other	Hospital Supply Service	
Packaging/Sterility	Other	
Quality Assurance	Sponsor	
Unknown	Overseas Advice	
Wear/Deterioration	Biomed Engineer	
	Para Medical	
	Patient/User	
	Dentist	
	Coroner	

Notes:

1. The problem causes are not mutually exclusive. For example, a material deficiency may have led to a mechanical malfunction or a biocompatibility problem.
2. Every report received by IRIS receives a risk analysis by the Scheme Coordinator and is discussed by a panel of technical and clinical professionals. In the case of reports that are "Not Investigated" the panel has made a decision that further investigation of the particular event is not necessary at that time. However, these reports are logged into the database for future reference and the trend of reports is monitored. In making their decision, the panel considers whether any similar reports have been received previously.